

263.

Accession number:20113414251284

Title:Optical fiber-coupled InGaAs-based terahertz time-domain spectroscopy system

Authors:Han, Sang-Pil (1); Ko, Hyunsung (1); Kim, Namje (1); Ryu, Han-Cheol (1); Wook Lee, Chul (1); Ahn Leem, Young (1); Lee, Donghun (1); Yong Jeon, Min (2); Kyu Noh, Sam (3); Sook Chun, Hyang (4); Hyun Park, Kyung (1)

Author affiliation:(1) THz Photonics Creative Research Center, Electronics and Telecommunications Research Institute (ETRI), Daejeon 305-700, Korea, Republic of; (2) Department of Physics, Chungnam National University, Daejeon 305-764, Korea, Republic of; (3) Global Research Lab., Nano Materials Evaluation Center, Korea Research Institute of Standards and Science (KRISS), Daejeon 305-340, Korea, Republic of; (4) Food Safety Research Division, Korea Food Research Institute, Sunghnam 463-746, Korea, Republic of

Corresponding author:Han, S.-P.(khp@etri.re.kr)

Source title:Optics Letters

Abbreviated source title:Opt. Lett.

Volume:36

Issue:16

Issue date:August 15, 2011

Publication year:2011

Pages:3094-3096

Language:English

ISSN:01469592

E-ISSN:15394794

CODEN:OPLEDP

Document type:Journal article (JA)

Publisher:Optical Society of America, 2010 Massachusetts Avenue NW, Washington, DC 20036-1023, United States

Abstract:The successful demonstration of an optical fiber-coupled terahertz time-domain spectroscopy (THz-TDS) system is described in this study. The terahertz output power of the emitter with two optical band rejection filters was 132nW, which is an improvement of 70% over the output power without any filters. This improvement is due to the suppression of an optical modulated signal that is reverse-generated when an alternating current bias exceeding a certain threshold is applied to the emitter. Under the optimal alignment conditions, the terahertz detector in a fiber-coupled THz-TDS system clearly measured water vapor dips in the free space.

Number of references:12