

190

Accession number:20114314458854

Title:Terahertz pulsed imaging in vivo: Measurements and processing methods

Authors:Parrott, Edward P. J. (1); Sy, Stanley M. Y. (1); Blu, Thierry (1); Wallace, Vincent P. (2); Pickwell-MacPherson, Emma (3)

Author affiliation:(1) Chinese University of Hong Kong, Department of Electronic Engineering, Shatin, Hong Kong; (2) University of Western Australia, School of Physics, Perth, WA 6009, Australia; (3) Hong Kong University of Science and Technology, Electronic and Computer Engineering, University of Science and Technology, Clearwater Bay, Hong Kong, Hong Kong

Corresponding author:Parrott, E.P.J.(e.pickwell.97@cantab.net)

Source title:Journal of Biomedical Optics

Abbreviated source title:J Biomed Opt

Volume:16

Issue:10

Issue date:October 2011

Publication year:2011

Article number:106010

Language:English

ISSN:10833668

E-ISSN:15602281

CODEN:JBOPFO

Document type:Journal article (JA)

Publisher:SPIE, P.O. Box 10, Bellingham, WA 98227-0010, United States

Abstract:This paper presents a number of data processing algorithms developed to improve the accuracy of results derived from datasets acquired by a recently designed terahertz handheld probe. These techniques include a baseline subtraction algorithm and a number of algorithms to extract the sample impulse response: double Gaussian inverse filtering, frequency-wavelet domain deconvolution, and sparse deconvolution. In vivo measurements of human skin are used as examples, and a comparison is made of the terahertz impulse response from a number of different skin positions. The algorithms presented enables both the spectroscopic and time domain properties of samples measured in reflection geometry to be better determined compared to previous calculation methods.

Number of references:51